


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION 4638</p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Qatar Armed Forces Calibration Center</h3> <p>Issue No: 001    Issue date: 17 June 2011</p>	
	<p><b>Calibration Center</b> PO Box 24129 Doha State of Qatar</p>	<p><b>Contact: QAF Calibration Center</b> Tel: +974-44616900/902/906 Fax: +974 - 44616901 / 921 EMail: ccqaf@yahoo.com Website:</p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
Electrical			
DC Voltage	1.018 V	4.1 ppm	
	10 V	1.3 ppm	
Resistance	10 $\Omega$	3.8 ppm	
	1 k $\Omega$	3.9 ppm	
Pressure			
Oil	2 MPa to 14 MPa	0.0070 %	
	14 MPa to 80 MPa	0.0045 %	
Mass			
Weight standard	1 kg	1.6 mg	
Mass	Nominal value		
	1 000 g	1.4 mg	
END			