

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 UKAS CALIBRATION 7540 Accredited to ISO/IEC 17025:2005	AeroCal Limited	
	Issue No: 001	Issue date: 14 November 2011
	Newhall Works Newhall Road Sheffield S9 2QL	Contact: Paul Adams Tel: +44 (0)114 230 0942 Fax: +44(0)114 230 0942 E-Mail: Website: www.aerocal.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Newhall Works Newhall Road Sheffield S9 2QL Local contact Paul Adams Tel: +44 (0)114 230 0942 Email: enquiries@erocal.co.uk	Head office	Lab

Site activities performed away from the location listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Electrical	Site



7540

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

AeroCal Limited

Issue No: 001 **Issue date:** 14 November 2011

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL				Lab & site
Generation and measurement				
DC Voltage	0 mV to 100 mV 100 mV to 10 V	0.055 mV 5.0 mV		
DC Current	0 mA to 20 mA	0.040 mA		
Electrical calibration of temperature simulators, indicators, controllers and recorders for the following sensors:				
Base metal thermocouples	- 200 °C to 0 °C 0 °C to 1370 °C	0.90 °C 0.60 °C	Including cold junction compensation	
Noble metal thermocouples	200 °C to 1760 °C	1.1 °C	Including cold junction compensation	
END				