


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>0248</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>WRc - NSF Ltd</p> <p>Issue No: 014 Issue date: 28 June 2007</p>	
	<p>Fern Close Pen-y-Fan Industrial Estate Oakdale Newport Gwent NP11 3EH</p>	<p>Contact: Mr Paul Davis Tel: +44 (0)1495 248454 Fax: +44 (0)1495 249234 E-Mail: info@wrcnsf.com Website:</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks
WATER Quantity of fluid passed	6 l to 5000 l (at flow rates of 0.1 l/min to 2000 l/min)	0.25%	Using standard start and finish method Calibration can be carried out up to 50 °C
END			