

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;">UKAS CALIBRATION 0831</p> <p style="text-align: center;">Accredited to ISO/IEC 17025:2005</p>	<h3 style="margin: 0;">Si-Plan Electronics Research Ltd</h3> <p style="margin: 0;">Issue No: 001 Issue date: 27 February 2006</p>	
	<p>Avenue Farm Industrial Estate Stratford-upon-Avon Warwickshire CV37 0HR</p>	<p>Contact: Mr T Stanberry-Flynn Tel: +44 (0)1789 205849 Fax: +44 (0)1789 415550 E-Mail: trevor.stanberry-flynn@si-plan.com Website: www.si-plan.com</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>Avenue Farm Industrial Estate Stratford-upon-Avon Warwickshire CV37 0HR</p>	<p>Mr T Stanberry-Flynn</p>	
<p>Any premises</p>		S



0831
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Si-Plan Electronics Research Ltd
Issue No: 001 Issue date: 27 February 2006

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
<p>FORCE</p> <p>UNIVERSAL TESTING MACHINES</p> <p>Verification and calibration of the force measuring system by force proving instruments in tension</p> <p>Verification and calibration of the force measuring system by force proving instruments in compression</p>	<p>From 0.5 kN up to 500 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2004</p> <p>From 0.5 kN up to 500 kN for Class 0.5, 1, 2 and 3 machines to BS EN ISO 7500-1:2004</p>	<p>0.22%</p> <p>0.22%</p>	<p>NOTES</p> <p>The uncertainties stated are for the application of the verification/calibration force and do not take into account the characteristics of the machine being verified/calibrated.</p> <p>In the place of the newton, verifications/ calibrations may be requested in the following units: ton-force (tonf), pound-force (lbf) or kilogram-force (kgf).</p>	S
END				