

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	<b>MIS Environmental Limited</b>	
	<b>Issue No: 037</b>	<b>Issue date: 15 November 2010</b>
	<b>Eden House</b> <b>Watling Street Industrial Estate</b> <b>Consett</b> <b>County Durham</b> <b>DH8 6TA</b>	<b>Contact: Mr A Nairn</b> <b>Tel: +44 (0)1207 500463</b> <b>Fax: +44 (0)1207 590240</b> <b>E-Mail: info@mis-environmental.co.uk</b> <b>Website: www.mis-environmental.co.uk</b>
<b>Testing performed by the Organisation at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Location Address</b> Eden House Watling Street Industrial Estate Consett County Durham DH8 6TA  <b>Local contact</b> Mr A Nairn  Tel: +44 (0)1207-500463 Fax: +44 (0)1207-590240 Email: info@mis-environmental.co.uk	Asbestos - Fibre counting	A
<b>Location Address</b> Suite 1 4 First Avenue Goole North Humberside DN14 6JQ  <b>Local contact</b> Mr A Nairn  Tel: +44 (0)1207-500463 Fax: +44 (0)1207-590240 Email: info@mis-environmental.co.uk	Asbestos - Fibre counting	B
<b>Location Address</b> Unit 2 Western Access Kestrel Road Trafford Park Manchester M17 1SF  <b>Local contact</b> Mr A Nairn  Tel: +44 (0)1207-500463 Fax: +44 (0)1207-590240 Email: info@mis-environmental.co.uk	Asbestos - Fibre counting	C

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
<b>Address</b> Any suitable premises including Domestic, Commercial and Industrial  <b>Local contact</b> Mr A Nairn	Asbestos - Fibre counting Asbestos - Sampling	D
Mobile Testing laboratories	Asbestos - Fibre counting Asbestos - Sampling	E



0244

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**MIS Environmental Limited**

**Issue No: 037 Issue date: 15 November 2010**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Methods for the Determination of Hazardous Substances in the series MDHS	
	Sampling of air for fibre counting	HSG 248:February 2005 Documented In-House Method Book 4, - Fibre Counting	D
	Fibre Counting	HSG 248:February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM) Documented In-House Method Book 4, - Fibre Counting	A, B, C, D, E
	4 Stage Clearance Process	HSG 248 February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM)	D
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for asbestos identification	Documented In-House Method Book 1, - Asbestos Sampling	D
ASBESTOS, CEMENT and PRODUCTS including Materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure, Book 2, using Stereo-microscopy, polarised light microscopy and dispersion staining	A, C
END			