


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	ADMAS Limited	
	Issue No: 019 Issue date: 30 January 2009	
	The Old Fellowship Hall 77 Storr Hill Wyke Bradford BD12 8PQ	Contact: Mrs L J Oakes Tel: +44 (0)1274 677111 Fax: +44 (0)1274 603838 E-Mail: office@admas.ltd.uk Website: www.admas.ltd.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address The Old Fellowship Hall 77 Storr Hill Wyke Bradford BD12 8PQ	Local contact Mrs L J Oakes Tel: +44 (0)1274 677 111 Fax: +44 (0)1274 603 838 Email: office@admas.ltd.uk Website: www.admas.ltd.uk	Support Functions Quality System Quality Audit Administration Sampling and Testing Health and Hygiene Occupational Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene Environmental Sampling	B
Mobile Laboratories	Health and Hygiene	C



0672

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

ADMAS Limited

Issue No: 019 Issue date: 30 January 2009

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	Documented In-House Method No OP12 based on HSG 248:February 2005	A, B, C
	Fibre counting	Documented In-House Method No OP05 based on HSG 248:February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method No OP27 based on HSG 248:February 2005	B, C
	Sampling of bulk materials for asbestos identification	Documented In-House Method No OP13 based on HSG 248:February 2005	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No OP06 based on HSG 248:February 2005, using stereo-microscopy, polarised light microscopy and dispersion staining	A
OCCUPATIONAL HYGIENE Dust and Particulates	<u>Analysis</u> Respirable and total inhalable airborne dusts and particulates	Documented In-House Method No OP22 (gravimetric method), based on Health and Safety Executive method MDHS 14/3:2000	A



0672
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

ADMAS Limited
Issue No: 019 Issue date: 30 January 2009

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
OCCUPATIONAL HYGIENE	<u>Sampling</u>		
Dust and Particulates	Sampling of ambient environments for subsequent analysis for airborne particulate matter respirable and total inhalable dust	Documented In-House Method No OP21 based on Health and Safety Executive method MDHS 14/3:2000	B
Gases and vapours including VOCs, acid halides, isocyanates, SO ₂ , ammonia, formaldehyde	Sampling of ambient environments for subsequent analysis for vapours and gases	Documented In-House Method No OP19 based on Health and Safety Executive methods MDHS 70:March 1990 MDHS 96:March 2000 MDHS 25/3:Jan1999 MDHS 6/3:March 1998 MDHS 47/2:June 1999 MDHS 84:June 1997 MDHS 101:Feb 2005	B
Airborne fibres	Sampling of ambient environments for subsequent airborne fibre concentration evaluation	Documented In-House Method No OP12 based on Health and Safety Executive methods HSG 248 and MDHS 59	B
END			