


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Adamson's Laboratory Services Ltd	
	Issue No: 017 Issue date: 17 June 2010	
	49 Lampits Hill Corringham Stanford-le-Hope Essex SS17 9AA	Contact: Miss J M Adamson Tel: +44 (0)1375 673279 Fax: +44 (0)1375 678059 E-Mail: jane@alsltd.com Website: www.alsltd.com

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 49 Lampits Hill Corringham Stanford-le-Hope Essex SS17 9AA	Local contact Miss J M Adamson	Asbestos A
Address Unit 6 Blott's Barn Brooks Road Raunds Northamptonshire NN9 6NS	Local contact Neil Palenski Tel: +44 (0)1933 461855 Fax: +44 (0)1933 461729 Email: npalenski@alstld.com	Asbestos B

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Asbestos	C



0675
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Adamson's Laboratory Services Ltd
Issue No: 017 Issue date: 17 June 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 using Documented In-House Procedure	A, B, C
	Sampling of air for fibre counting	HSG 248:February 2005 using Documented In-House Procedure	C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 using Documented In-House Procedure	C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure ASTP 2 using stereo-microscopy, polarised light microscopy and dispersion staining	A, B
	Sampling of bulk materials for asbestos identification	Documented In-House Method ASTP 2 in the Procedures Manual	C
END			