


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	ACS Physical Risk Control Ltd	
	Issue No: 017 Issue date: 28 September 2009	
	14 Claremont Centre Durham Street Glasgow G41 1BS	Contact: Prof R J Willey Tel: +44 (0)141-427 5171 Fax: +44 (0)141-427 2722 E-Mail: roger@acs-env.com Website: www.acs-env.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 14 Claremont Centre Durham Street Glasgow G41 1BS	Local contact Prof R J Willey Tel: +44 (0)141-427 5171 Fax: +44 (0)141-427 2722 Email: roger@acs-env.com Website: www.acs-env.com	Health and Hygiene A
Address Unit 7A Damery Works Woodford Berkeley Gloucester GL13 9JR	Local contact Ms E Willey Tel: 07811 387 026 Fax: +44 (0) 145 426 9501 Email: emma@acs-env.com Website: www.acs-env.com	Health and Hygiene B

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	C



0676

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

ACS Physical Risk Control Ltd

Issue No:017 Issue date: 28 September 2009

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure) Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B, C
	Sampling of air for fibre counting	HSG 248:February 2005 by Documented In-House Method No ACS/T5, 6, 7, 8, 9	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Method No ACS/T2 using stereo-microscopy, polarised light microscopy and dispersion staining	A,
	Sampling of bulk materials for asbestos identification	Documented In-House Method No ACS/T1, 2, 3, 5, 6, 9	C
END			