


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>Accredited to ISO/IEC 17025:2005</b>	<b>Babcock Marine [Rosyth] Ltd trading as Babcock Scientific</b>	
	<b>Issue No: 024    Issue date: 21 December 2010</b>	
	<b>Babcock Support Services Ltd Great Michael Road Rosyth Royal Dockyard Rosyth Fife KY11 2YD</b>	<b>Contact: Mr S Neithercut Tel: +44 (0)1383 424100 Fax: +44 (0)1383 422699 E-Mail: Stewart.Neithercut@babcock.co.uk Website: www.babcock.co.uk</b>

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Babcock Marine [Rosyth] Ltd trading as Babcock Scientific Great Michael Road Rosyth Royal Dockyard Rosyth Fife KY11 2YD	<b>Local contact</b> Mr S Neithercut  Tel: +44 (0)1383 424100 Fax: +44 (0)1383 422699 Email: Stewart.Neithercut@babcock.co.uk Website: www.babcock.co.uk	Health and Hygiene  A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



0757

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Babcock Marine [Rosyth] Ltd**  
trading as  
**Babcock Scientific**

**Issue No: 024 Issue date: 21 December 2010**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	B
	Fibre counting	HSG 248:February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	B
	Sampling of bulk materials for asbestos identification	HSG 248:February 2005 (Documented In-House Procedure)	B
WATERS Trade effluent (to controlled water) and process water	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 (Documented In-House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining)	A
	<u>Radiation Tests</u> Qualitative and quantitative analysis of gamma-emitting radionuclides: 80 - 2000 keV (Quantitative) 46 - 2000 keV (Qualitative)	Documented In-House Method using Computerised Gamma-Ray Spectrometry RCI Manual, Vol I, Method 6.24	A
END			