


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	PTL Occupational Hygiene Consultants Ltd	
	Issue No: 012	Issue date: 16 December 2008
	1 Church Walk Stapleford Nottingham NG9 8DE	Contact: Mr T Lines Tel: +44 (0)115 939 7581 Fax: +44 (0)115 939 2686 E-Mail: terry@ptlhygiene.co.uk Website: www.ptlhygiene.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 1 Church Walk Stapleford Nottingham NG9 8DE	Local contact Mr T Lines	Health and Hygiene
		A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Mobile Laboratories	Local contact Mr T Lines	Health and Hygiene
Premises including domestic, commercial and industrial	Local contact Mr T Lines	Health and Hygiene
		B
		C



1590
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

PTL Occupational Hygiene Consultants Ltd
Issue No: 012 Issue date: 16 December 2008

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG248: February 2005 by Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B, C
	Sampling of air for fibre counting	HSG248: February 2005 (Documented In-House Procedure)	B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG248: February 2005 (Documented In-House Procedure)	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248: February 2005 by Documented In-House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for asbestos identification	Documented In-House Method	B, C

END