


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Antech Calibration Services	
	Issue No: 005 Issue date: 04 May 2007	
	Hewett Road Gapton Hall Industrial Estate Great Yarmouth Norfolk NR31 0NN	Contact: Mr J Gunn Tel: +44 (0)1493-440600 Fax: +44 (0)1493-440600 E-Mail: sales@antech.org.uk Website: www.antech.org.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Hewett Road Gapton Hall Industrial Estate Great Yarmouth Norfolk NR31 0NN	Local contact Mr J Gunn	Pressure relief valve testing Bonded Pressure seals Permanent

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer premises - any	Pressure relief valve testing	Site



1828

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Antech Calibration Services
Issue No: 05 Issue date: 04 May 2007

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Spring operated pressure relief valves (Conventional and balanced bellows)	Pre-inspection Set pressure	BS 6759 : Part 1 : 1984 Section 6 Clause19 BS 6759 : Part 2 : 1984 Section 6 Clause12	Laboratory and site
Pilot-operated pressure relief valves (Both pilot only and combination tested)	Seat tightness Bellows pressure Body pressure 'Blowdown' Testing on water up to 700 bar Testing on gas up to 350 bar	BS 6759 : Part 3 : 1984 Section 3 Clause 6.9 & 6.10 Section 4 Clause 6.17 & 10.5 API standard 527, 3 rd edition July 1991 Section 2,4 and 5 Valves can be tested to the set pressure tolerances set out in: ASME I and VIII Code Division 1 Pressure Vessels – 1977	Laboratory and site
Bonded Pressure seals	Burst pressure test Testing on Oil up to 2000 bar	In-house procedure 22000	Laboratory
END			