


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>Accredited to ISO/IEC 17025:2005</b>	<b>AWE PLC</b>	
	<b>Issue No: 010    Issue date: 11 June 2010</b>	
	<b>Building A14.1 Aldermaston Reading Berks RG7 4PR</b>	<b>Contact: No Commercial Enquiries Tel: Fax: E-Mail: Website:</b>

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> Building A14.1 Aldermaston Reading Berks RG7 4PR	<b>Local contact</b> No Commercial Enquiries  Tel: Fax: Email:	Health and Hygiene	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including commercial, domestic and industrial	Health and Hygiene	B



1854

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**AWE PLC**

**Issue No: 010    Issue date: 11 June 2010**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	Membrane Filter Method using Phase Contrast Microscopy (PCM)	A
	Sampling of air for fibre counting	Documented In-House Method WI 204	B
ASBESTOS IN BULK MATERIALS including materials and products suspected on containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure) Membrane Filter method using Phase Contrast Microscopy (PCM)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Method WI 206 using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for asbestos identification	Documented In-House Method WI 205	B
END			