

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>Accredited to ISO/IEC 17025:2005</b>	<b>Abu Dhabi Food Control Authority (ADFCA)</b>	
	Issue No: 020 Issue date: 23 April 2012	
	<b>Delma Street No 13 PO Box 52150 Abu Dhabi United Arab Emirates</b>	<b>Contact: Mr Tariq MN Al Hammadi Tel: +9712 495 4494 Fax: +9712 446 3729 E-Mail: tariq.alhammadi@adfca.ae Website: www.adfca.ae</b>

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code	
<b>Address</b> Delma Street no 13 PO Box 52150 Abu Dhabi United Arab Emirates	<b>Local contact</b> Mr Badr Al Shehhi Quality Manager	Chemical and Microbiological Testing Quality Management systems co-ordination including document control, PT scheme participation, internal audits, purchasing and supplies	A
<b>Address</b> Al Ain Branch Government District Court Road PO Box 66066 Al Ain United Arab Emirates	<b>Local contact</b> Mr Badr Al Shehhi Quality Manager	Chemical and Microbiological Testing	B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**

**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS, LEACHATES, POLLUTANTS - EFFLUENTS: LIQUID SEDIMENTS, SOILS, WATERS	<u>Chemical Tests</u>	Documented In-House Methods referenced by Standard Operating Procedure (SOP) number based on/ incorporating standard methods as stated	
	<u>Radiochemical Tests</u>		
WATER	Cs-137	LAD-DPM-RL-01 by Gamma Spectroscopy	A
Raw and potable	Gross radium alpha activity	LAD-DPM-RL-02 by precipitation of radium followed by alpha emission counting	A
Drinking/potable water	Gross alpha and gross beta radioactivity in water	LAD-DPM-RL-04 using gas-flow low-level planchet counting system	A
FATS and OILS	<u>Chemical Tests</u>		
	Acid value and free fatty acid	LAB-DPM-CHM-01 based on AOAC International, 18th Edition 2005	A, B
	Colour	LAD-DPM-CHM-02 based on Pearson's composition and analysis of foods, 9th Edition 1991	A
	Preparation of Methyl esters of Fatty acid	LAD-DPM-CHM-04 by saponification followed by methylation based on AOAC International, 18th Edition 2005	A
	Fatty acid composition - by capillary GC cis cis and trans isomers: Caproic C 6:0 Heptanoic C 7:0 Caprylic C 8:0 Nonanoic C 9:0	LAD-DPM-CHM-05 based on AOAC 963.22 and AOCS Ce-1c-89 using capillary column GC	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**

**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FATS and OILS (cont'd)	<p><u>Chemical Tests</u> (cont'd)</p> <p>Fatty acid composition - by capillary GC cis cis and trans isomers: (cont'd)</p> <p>Capric C10:0 Undecanoic C11:0 Lauric C12:0 Tridecanoic C13:0 Myristic C14:0 Myristoleic C14:1 Pentadecanoic C15:0 Pentadecenoic C15:1 Palmitic C16:0 Palmitoleic C16:1 Heptadecanoic C17:0 Heptadecenoic C17:1 Stearic C18:0 Oleic trans C18:1 Oleic cis-9 C18:1 Oleic cis-7 C18:1 Linoleic tr, tr, C18:2 Linoleic tr, cis, C18:2 Linoleic cis, tr, C18:2 Linoleic cis,cis, C18:2 Tr, tr, tr, Linolenic C18:3 Tr, tr, cis, Linolenic C18:3 Tr, cis, tr, Linolenic C18:3 Cis, tr, tr, Linolenic C18:3 Cis, cis, tr, Linolenic C18:3 Cis, tr, cis, Linolenic C18:3 Tr, cis, cis, Linolenic C18:3 Cis, cis, cis, Linolenic C18:3 Nonadecanoic C19:0 Arachidonic C20:0 Gadoleic C20:1 Arachidonic C20:4 Henicosanoic C21:0 Behenic C22:0 Erucic C22:1 Dosanoic C23:0 Lignoceric C24:0 Nervonic C24:1 Behenic C22:0 Erucic C22:1 Dosanoic C23:0</p>	LAD-DPM-CHM-05 based on AOAC 963.22 and AOCS Ce-1c-89 using capillary column GC (cont'd)	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FATS and OILS (cont'd)	<u>Chemical Tests</u> (cont'd)		
	Fatty acid composition - by capillary GC cis cis and trans isomers: (cont'd)	LAD-DPM-CHM-05 based on AOAC 963.22 and AOCS Ce-1c-89 using capillary column GC (cont'd)	A
	Lignoceric C24:0 Nervonic C24:1		
	Iodine Value	LAD-DPM-CHM-01 based on AOAC International, 18th Edition 2005 LAA-DPM-CHM-05 based on AOAC International 18th Edition: 2005	A, B
	Peroxide Value	LAB-DPM-CHM-04 by titration based on AOAC International, 18th Edition 2005	A, B
	Rancidity (Kreis Test)	LAD-DPM-CHM-03 based on Pearson's composition and analysis of foods, 9th Edition 1991	A
	Refractive Index	LAB-DPM-CHM-05 based on AOAC International, 18th Edition 2005 using digital refractometer	A B
	Saponification Value	LAB-DPM-CHM-02 by titration	A, B
Olive Oil	Specific Extinction	LAA-DPM-CHM-04 using (UV/Vis) Spectrophotometry	B
	Specific Extinction	LAD-DPM-CHM-07 using UV/Vis spectrophotometry	A
FOODS	<u>Chemical Tests</u>		
Dates	Pesticide Residues Analysis: Alpha-Cypermethrin Benfuracarb Bromoprylate Cadusafos Carbosulfan Chlorfenvinphos Chlorpyrifos Ethyl Chlorpyrifos Methyl Diazinon	LAD-DPM-CHPR-04 using solid phase extraction (SPE) and GC MSD	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)	<u>Chemical Tests</u> (cont'd)		
Dates (cont'd)	Pesticide Residues Analysis: (cont'd)  Dichlorovos Endosulfan 1 p,p-DDE Endosulfan 2 o,p-DDT p,p-DDD Endosulfan sulphate Fenprothrin Gamma HCH L-Cyhalothrin Malthion Methidathion Phosalone Pyridabin Triazophos Pirimifos methyl	LAD-DPM-CHPR-04 using solid phase extraction (SPE) and GC MSD (cont'd)	A
Beverages, Fruit Juices and Drinks	Acesulfame K Aspartame Benzoic acid and salts Caffeine Saccharin Sorbic acid and salts	LAD-DPM-CHAV-05 by High Performance Liquid Chromatography (HPLC)	A
Coffee (Green)	Ochratoxin A	LAD-DPM-CHSP-03 by Fluorimetry	A
Foods and Beverages	Alcohol	LAD-DPM-CHM-06;TM69 based on AOAC 973.23 by Gas Chromatography	A
Foodstuffs, General	Sweeteners & Preservatives: Acesulfame K Saccharin	LAA-DPM-CHAV-04 by High Performance Liquid Chromatography (HPLC)	B
	Moisture and Volatile Matter	LAA-DPM-CHN-01 based on Pearson's composition and analysis of foods, 8th Edition using hot air oven  LAB-DPM-CHNA-03 using Halogen Moisture Analyser	B  B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**

**Issue No: 020 Issue date: 23 April 2012**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)	<u>Chemical Tests</u> (cont'd)		
Drinks, Juices and Jam	Preservatives: Sodium benzoate Caffeine Potassium sorbate	LAA-DPM-CHAV-01 by High Performance Liquid Chromatography (HPLC)	B
	Artificial Food Colours	LAA-DPM-CHAV-02 by extraction and Chromatography	B
Honey		LAA-DPM-CHAV-03 Extraction of colour	B
	Hydroxymethylfurfural (HMF)	LAB-DPM-CHM-03 by Ultra-Violet Spectrophotometry (UV)	A, B
	Moisture	LAA-DPM-CHNA-07 using Digital Refractometer	B
	Pesticide multi-residue screen: Alphacypermethrin Azoxystrobin Bromophos ethyl Bromophos methyl Chlorpyrifos Chlorpyrifos methyl Coumaphos Cyanophenphos Deltamethrin Diazinon Dichlofluanid Ethion Febuconazole Fenamiphos Fenarimol Fenthion Flusilazole Isenophos Methyl Kresoxim Methyl Malathion Methiocarb Pendimethalin Phosalone Procymidone Quinalphos Tecnazene Tetradifon	LAD-DPM-CHPR-07 by GCMS	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)	<u>Chemical Tests</u> (cont'd)		
Honey (cont'd)	Pesticide multi-residue screen: (cont'd)  Tolyfluanide Triadumefon Trifloxystrobin	LAD-DPM-CHPR-07 by GCMS (cont'd)	A
Foods and Beverages	Sugars covering: Dextrose Fructose Maltose Lactose Sucrose	LAD-DPM-CHNA-17 by HPLC	A
Meat	Identification of pork and pork derivatives	LAD-DPM-CHSP-05 using ELISA	A
Meat	Identification of pork and pork derivatives	LAA-DPM-CHSP-03 using ELISA	B
Milk, Raw and Pasteurised	Fat Lactose Protein Total solids and solids - non-fat (SNF)	LAB-DPM-CHNA-01 using Near Infra-Red Spectroscopy using Milkoscan FT120	A
Milk and Dairy Products	Acidity	LAD-DPM-CHNA-02 by titration	A
	Ash	LAD-DPM-CHNA-07 by ignition at 550 °C	A
	Fat	LAD-DPM-CHNA-01 by Gerber method	A
	Starch (qualitative)	LAD-DPM-CHNA-04 using iodine blue colour indication	A
	Total Solids	LAD-DPM-CHNA-06 by loss of volatile material at 102 °C	A
Milk Powder and Date Seed	Nitrogen and protein	LAA-DPM-CHNA-02 by automated Kjeldahl determination	B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)	<u>Chemical Tests</u> (cont'd)		
Milk Powder and Orange Juice	Titrateable Acidity	LAA-DPM-CHNA-05 titration	B
Milk, Milk Powder, Yogurt and Yogurt Drink	Melamine	LAD-DPM-CHSP-04 using enzyme-linked immunosorbant assay	A
Grains and Feeds	Total aflatoxin	LAD-DPM-CHSP-02 using VICAM Fluorometer	A
		LAA-DPM-CHSP-02 using VICAM Fluorometer	B
Nuts	Total aflatoxin	LAD-DPM-CHSP-01 using Fluorometer	A
		LAA-DPM-CHSP-01 using Fluorometer	B
Food and Food products except butter	Sodium chloride	LAD-DPM-CHNA-03 by Mohr Method	A
		LAA-DPM-CHNA-04 by Mohr Method	B
Fruit and Vegetables	Pesticide multi-residue screen: Alpha Cypermethrin Aldrin Azinophos methyl Benalaxyl Benfuracarb benzoximate Bifenthrin Bitranol Bromoprylate Buprofezin Cadusafos Captan Carbofuran Carbosulfan Chlorfenvinfos Chloroprotham Chlorpyrifos-ethyl Chlorpyrifos-methyl Cis-Permethrin	LAD-DPM-CHPR-02 using Gas Chromatography with ECD, NPD and MS detectors	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**

**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd) Fruit and Vegetables (cont'd)	<u>Chemical Tests</u> (cont'd) Pesticide multi-residue screen: (cont'd) DDD- pp DDE -pp DDT- o,p DDT- p,p Delta HCH Diazinon Dichloran Dichlorvos Dicofol Dieldrin Difenthrion Dimethoate Diphenylamine Endosulfan-1 Endosulfan-II Endosulfan-sulphate Endrin Endrin aldehyde Ethion Ethoprophos Fenarimol Fenvalerate Flamprop Methyl Folpet Iprodione kresoxim-methyl lambda cyhalothrin Malathion Metalaxyl Methoxychlor Monocrotophos Myclobutanil Oxadixyl Parathion Ethyl Parathion methyl Penconazole Pendimethalin Phenyl Phenol-2 Phosalone Phosmet Procymidone Propham	LAD-DPM-CHPR-02 using Gas Chromatography with ECD, NPD and MS detectors (cont'd)	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd) Fruit and Vegetables (cont'd)	<u>Chemical Tests</u> (cont'd) Pesticide multi-residue screen: (cont'd) Pyridabin Pyrimethanil Quentozene Quinalphos Tecnazene Tolclophos methyl Tolyfluanid Trans-Permethrin Triazophos Trifluralin Tryfloxistrobin Fenitrothion Fenpropathrin Buprimate Fenitrothion Fenpropathrin Methidathion Primicarb Vinclozolin Furalaxyl Gamma HCH Heptachlor Heptenphos Hexachlorobenzene	LAD-DPM-CHPR-02 using Gas Chromatography with ECD, NPD and MS detectors (cont'd)	A
Fruit and Vegetables (High water content)	Pesticide multi-residue screen: 2-Phenyl-phenol 4,4'-DDD 4,4-DDE 4,4-DDT Acephate Acrinathrin Aldrin Alpha-Endosulfan Alpha-HCH Atrazine Azinphos-methyl Azoxystrobin Benalexyl Beta-endosulfan Beta-HCH Bifenthrin	LAA-DPM-CHPR-02 using GCMS	B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**

**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)  Fruit and Vegetables (High water content) (cont'd)	<u>Chemical Tests</u> (cont'd)  Pesticide multi-residue screen: (cont'd)  Bitertanol Bitertanol BD Bromophos-methyl Bromopropylate Bupimate Buprofezin Cadusafos Carbaryl Carbofuran Chlorfenvinfos Chlorobufam Chloropyrifos-methyl Chlorpropham Chlorpyrifos Cis-chlordane Cypermethrin Delta-HCH Deltamethrin Diazinon Dichlofenthion Dichlofluanid Dicloran Diclorvos Dicofol Dicrotophos Dieldrin Difenoconazole Dimethoate Diniconazole Diphenylamine Endosulfan sulfate Endrin Endrin Aldehyde Ethalfuralin Ethion Ethoprophos Fenarimol Fenbuconazole Fenhexamid Fenitrothion Fenthion Fenvalerate	LAA-DPM-CHPR-02 using GCMS (cont'd)	B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)  Fruit and Vegetables (High water content) (cont'd)	<u>Chemical Tests</u> (cont'd)  Pesticide multi-residue screen: (cont'd)  Fipronil Flamprop-methyl Fludioxonil Flusilazole Folpet Furalaxyl Gamma-HCH Heptachlor Heptachlor-exo-epoxide- Heptonophos Hexaconazole Iprodione Isofenphos-methyl Kresoxim-methyl Lambda-Cyhalothrin Malathion Malaxon Mecarbam Metalaxyl Methidathion Methiocarb Myclobutanil Oxadixyl Paraoxon-methyl Parathion-ethyl Parathion-methyl Penconazole Pendimethalin Permethrin Phorate Phosalone Phosmet Phosphamidon Pirimiphos-methyl Primicarb Procymidone Profenofos Prometon Propiconazole Propyzamide Quinalphos Quintozene	LAA-DPM-CHPR-02 using GCMS (cont'd)	B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)	<u>Chemical Tests</u> (cont'd)		
Fruit and Vegetables (High water content) (cont'd)	Pesticide multi-residue screen: (cont'd)  Tebuconazole Tecnazene Tetraconazole Tetradifon Tetramethrin Tolclofos-methyl Tolyfluanid Triadimefon Triadimenol Trifloxystrobin Trifluralin Vinclozolin	LAA-DPM-CHPR-02 using GCMS (cont'd)	B
FOODS, General	Artificial colours (qualitative identification)	LAD-DPM-CHAV-03 using paper chromatography	A
	Crude Fibre	LAD-DPM-CHNA-14 using Fibretech system	A
	Sudan Red dyes	LAD-DPM-CHAV-01 by HPLC	A
ANIMAL FEEDINGSTUFFS FOODS, General	Ash	LAD-DPM-CHNA-10 by ignition at 550 °C	A
	Crude Fibre	LAA-DPM-CHNA-10 using Fibretech system	B
	Fat	LAD-DPM-CHNA-12 by direct Soxhlet extraction	A
	Fat Content	LAD-DPM-CHNA-15 by acid hydrolysis and Soxhlet extraction	A
	Moisture	LAD-DPM-CHNA-11 by loss of volatile matter at 102 °C	A
	Nitrogen and crude protein	LAD-DPM-CHNA-13 by automated Kjeldahl determination	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**

**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOOD	<u>Chemical Tests</u> (cont'd) Aluminium Antimony Arsenic Barium Beryllium Cadmium Calcium Chromium Cobalt Copper Iron Lead Magnesium Manganese Mercury Molybdenum Nickel Phosphorus Potassium Selenium Silver Sodium Thallium Thorium Uranium Vanadium Zinc	LAD-DPM-CHI-16 using Inductively Coupled Plasma with Mass Spectrometry (ICP-MS)	A
FOOD	Copper Iron Magnesium Potassium Sodium Zinc	LAA-DPM-CHI-07 using Flame Atomic Absorption Spectrometry	B
	Arsenic Cadmium Lead	LAA-DPM-CHI-08 using Zeeman Atomic Absorption Spectrometry, Graphite Tube Atomizer	B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**

**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code	
FOODS  Dairy Products	<u>Microbiological Tests</u>			
	Enumeration of:			
	Coliforms and Escherichia coli	LAB-DPM-MICRO-07 using Most Probable Number technique	A	
	E Coli O157	LAB-DPM-MICRO-21 using selective enrichment. Confirmation with SAS visual assay and VITEK2.	A	
	Coagulase positive staphylococci including Staphylococcus aureus	LAB-DPM-MICRO-08 based on ISO 6888-1:1999, confirmation using VITEK2 and staph slidex	A	
	Total colony count	LAB-DPM-MICRO-10 based on ISO 4833:2003 at 37°C for 48h	A	
	Infant and Dietetic Foods	<i>Bacillus cereus</i>	LAB-DPM-MICRO-14 using spread plate technique, confirmation using VITEK2	A
		<i>Clostridium perfringens</i>	LAB-DPM-MICRO-06 using pour plate technique based on ISO 7937:2004, confirmation using VITEK2	A
		Total colony count (37 °C)	LAB-DPM-MICRO-10 based on ISO 4833:2003 at 37°C for 48h	A
		Detection of:		
Meat and Meat Products	<i>Cronobacter sakazakii</i> ( <i>Enterobacter sakazakii</i> )	LAB-DPM-MICRO-19 based on ISO22964:2006	A	
	Enumeration of:			
	<i>Bacillus cereus</i>	LAB-DPM-MICRO-14 using spread plate technique, confirmation using VITEK2	A	
	Coagulase positive staphylococci including <i>Staphylococcus aureus</i>	LAB-DPM-MICRO-08 SO 6888-1: 1999 based on, confirmation using VITEK2 and staph slidex	A	



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)	<u>Microbiological Tests</u> (cont'd)		
Foods, General	Enumeration of:		
	Total colony count (37 °C)	LAB-DPM-MICRO-10 based on ISO 4833:2003 at 37°C for 48h	A, B
	<i>Bacillus cereus</i>	LAB-DPM-MICRO-14 using spread plate technique on selective agar. Confirmation using VITEK2	B
	<i>Clostridium perfringens</i>	LAB-DPM-MICRO-06	B
	Coliforms and <i>Escherichia coli</i>	LAB-DPM-MICRO-07 using Most Probable Number technique	A
		LAB-DPM-MICRO-07 using Most Probable Number technique. E. coli confirmation using VITEK2	B
	Coagulase positive staphylococci including <i>Staphylococcus aureus</i>	LAB-DPM-MICRO-08 based on ISO 6888-1:1999. Confirmation using VITEK2 and staph slidex	B
	Detection of:		
	<i>E coli</i> O157	LAB-DPM-MICRO-21 using selective enrichment. Confirmation with SAS visual assay and VITEK2	A, B
	<i>Listeria monocytogenes</i>	LAB-DPM-MICRO-04 using selective enrichment culture, confirmation using VITEK2	A
	<i>Salmonella</i> spp	LAB-DPM-MICRO-03 based on ISO6579:2002+A1:2007 Confirmation using serology, Biomerieux API and/or VITEK2	A, B
	<i>Vibrio</i> spp	LAB-DPM-MICRO-13 using selective enrichment culture. Confirmation using Biomerieux APland/or VITEK 2	A
Cheese	<i>Listeria monocytogenes</i>	LAB-DPM-MICRO-04 using selective enrichment culture. Confirmation using VITEK2	B



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FOODS (cont'd)	<u>Microbiological Tests</u> (cont'd)		
Foods, General (cont'd)	Detection of: (cont'd)		
Fish, Fish products and Shellfish	<i>Vibrio</i> spp	LAB-DPM-MICRO-13 using selective enrichment culture. Confirmation using VITEK2	B
WATERS	<u>Chemical Tests</u>		
Drinking Water	Aluminium Antimony Arsenic Barium Beryllium Cadmium Chromium Cobalt Copper Lead Manganese Molybdenum Nickel Selenium Silver Thallium Uranium Vanadium Zinc	LAA-DPM-CHI-06 using Inductively Coupled Plasma with Mass Spectrometry (ICP-MS)	B
	Aluminium Antimony Arsenic Barium Beryllium Cadmium Calcium Chromium Cobalt Copper Iron Lead Magnesium Manganese Mercury Molybdenum Nickel	LAD-DPM-CHI-16 using Inductively Coupled Plasma with Mass Spectrometry (ICP-MS)	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)		
	Continued from Page 19		
Drinking Water (cont'd)	Phosphorus Potassium Selenium Silver Sodium Thallium Thorium Uranium Vanadium Zinc		
	Arsenic Cadmium Lead	LAA-DPM-CHI-08 using Zeeman Atomic Absorption Spectrometry, Graphite Tube Atomizer	B
	Total Hardness and Calcium Hardness as CaCO <sub>3</sub>	LAA-DPM-CHI-03 by titration	B
	Conductivity and Total dissolved solid	LAA-DPM-CHI-01 using conductivity meter	B
	Copper Iron Magnesium Potassium Sodium Zinc	LAA-DPM-CHI-07 using Flame Atomic Absorption Spectrometry	B
	Determination of disinfection products: ClO <sub>2</sub> , ClO <sub>3</sub> and BrO <sub>3</sub> ions	LAD-DPM-CHI-14 ion chromatography using IC90	A
	Determination of disinfection products: ClO <sub>2</sub> , ClO <sub>3</sub> , Br, and BrO <sub>3</sub> ions	LAA-DPM-CHI-04 ion chromatography using IC3000	B
	Determination of inorganic ions: Br, Cl, F, NO <sub>2</sub> , NO <sub>3</sub> , PO <sub>4</sub> and SO <sub>4</sub>	LAD-DPM-CHI-15 ion chromatography using IC3000	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (cont'd)	<u>Chemical Tests</u> (cont'd)		
Drinking Water (cont'd)	Determination of inorganic ions: Cl, F, NO <sub>2</sub> , NO <sub>3</sub> , PO <sub>4</sub> and SO <sub>4</sub>	LAA-DPM-CHI-05 ion chromatography using IC3000	B
	pH	LAA-DPM-CHI-02 using pH meter	B
Raw, potable	Bromide Chloride Fluoride Nitrate Phosphate Sulphate	LAD-DPM-CHI-02 based on AOAC Method 993.30 using Ion-Chromatography (IC)	A
	Organochlorine pesticides: aldrin 4,4'-DDD 4,4'-DDE 4,4'-DDT Dieldrin endosulphan (as sum of endosulphan I endosulphan II, endosulphan sulphate) endrin (as sum of endrin, endrin aldehyde, endrin ketone) alpha-HCH beta-HCH delta-HCH gamma-HCH heptachlor heptachlor-epoxide methoxychlor	LAD-DPM-CHPR-01 using Solid Phase Extraction (SPE) and Gas Chromatography and confirmation of detected residues by Gas Chromatography Mass Spectrometry	A



2028  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Abu Dhabi Food Control Authority (ADFCA)**  
**Issue No: 020 Issue date: 23 April 2012**

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATER  Raw and Potable Bottled Drinking Water	<u>Microbiological Tests</u>		
	Enumeration of:		
	Total Colony Count (22 °C and 37 °C)	LAB-DPM-MICRO-10 using spread plate technique based on The Microbiology of Drinking Water 2002:Part 7	A
	Coliforms	LAB-DPM-MICRO-22 using membrane filtration, based on The Microbiology of Drinking Water. Part 4:2009	A, B
	<i>E. coli</i>	LAB-DPM-MICRO-22 using membrane filtration, based on The Microbiology of Drinking Water. Part 4:2009	A, B
Enterococci	LAB-DPM-MICRO-29 based on ISO 7899-2:2000	A, B	
<i>Pseudomonas aeruginosa</i>	LAB-DPM-MICRO-30 based on ISO 16266:2006	A, B	
END			