


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	JB-Monitor Ltd	
	Issue No: 011 Issue date: 18 January 2011	
	11 Wynnstay Road Colwyn Bay Conwy LL29 8NB	Contact: Mr J J Bullock Tel: +44 (0)1492 536444 Fax: +44 (0)1492 536444 E-Mail: info@jbmonitor.co.uk Website: www.jbmonitor.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address 11 Wynnstay Road Colwyn Bay Conwy LL29 8NB	Local contact Mr J J Bullock Tel: +44 (0)1492 536444 Fax: +44 (0)1492 536444 Email: info@jbmonitor.co.uk	Asbestos	Permanent

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Domestic, Industrial, Government, Commercial premises	Contact: Mr J J Bullock	Asbestos	Site



2075

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

JB-Monitor Ltd

Issue No: 011 Issue date: 18 January 2011

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 (In house method No 2), Membrane Filter Method using Phase Contrast Microscopy (PCM)	Permanent & site
	Sampling of air for fibre counting	HSG 248:February 2005 (In house method No 1)	Site
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (In-House Method)	Site
	Sampling of bulk materials for asbestos identification	HSG 248:February 2005 (In-House Method No 3)	Site
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 (In-House Method No 4) using stereo-microscopy, polarised light microscopy and dispersion staining	Permanent
END			