


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Ethos Environmental Ltd	
	Issue No: 008 Issue date: 15 March 2010	
	Unit 16 Dumbryden Industrial Estate 32 Dumbryden Road Edinburgh EH14 2AB	Contact: Dr B Gardner Tel: +44 (0)131 453 5111 Fax: +44 (0)131 453 6111 E-Mail: info@ask-ethos.co.uk Website: www.ask-ethos.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 16 Dumbryden Industrial Estate 32 Dumbryden Road Edinburgh EH14 2AB Local contact Dr B Gardner	Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



2146
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Ethos Environmental Ltd
Issue No: 008 Issue date: 15 March 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre Counting	Documented In-House Method (TP003) based on Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure TP002) based on Membrane Filter Method using Phase Contrast Microscopy (PCM)	B
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure TP002) based on Membrane Filter Method using Phase Contrast Microscopy (PCM)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Method TP004 using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for asbestos identification	Documented In-House method TP005	B
END			