

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	<h3>DMW Environmental Safety</h3> <p>Issue No: 014 Issue date: 22 January 2010</p>	
	<p>Unit 4 Calibre Industrial Park Laches Close Four Ashes Wolverhampton WV10 7DZ</p>	<p>Contact: Mr D Redfern Tel: +44 (0)1902 791565 Fax: +44 (0)1902 791595 E-Mail: DavidRedfern@dmwsafety.co.uk Website: www.dmwsafety.co.uk</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Unit 4 Calibre Industrial Park Laches Close Four Ashes Wolverhampton WV10 7DZ</p> <p>Local contact Mr D Redfern Tel: +44 (0)1902 791565 Fax: +44 (0)1902 791565 Email: DavidRedfern@dmwsafety.co.uk Website: www.dmwsafety.co.uk</p>	Health and Hygiene	A
<p>Address DMW Environmental Safety E-on/UK Power Station Ratcliffe-on-Soar Nr Nottingham NG11 0EE</p> <p>Local contact Mr D Redfern Tel: +44 (0)1902 791565 Fax: +44 (0)1902 791565 Email: DavidRedfern@dmwsafety.co.uk Website: www.dmwsafety.co.uk</p>	Health and Hygiene	B

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>DMW Environmental Safety E-on/UK Power Station Ratcliffe-on-Soar Nr Nottingham NG11 0EE</p>	Asbestos	B
<p>Premises including domestic, commercial and industrial</p>	Health and Hygiene	C
<p>Mobile Laboratory</p>	Asbestos	D



2167

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

DMW Environmental Safety
Issue No: 014 Issue date: 22 January 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure) Membrane Filter Method by Phase Contrast Microscopy (PCM), using SOP 9	A, B, C
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure) using SOP 09	B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure) Membrane Filter Method using Phase Contrast Microscopy (PCM) using SOP 9	B, C
	Sampling of bulk materials for asbestos identification	Documented In-house Method No SOP 10	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by stereo microscopy, polarised light microscopy and dispersion staining using documented In-house Methods: SOP10, SOP14, SOP15	A, B, D
END			