


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>Accredited to ISO/IEC 17025:2005</b>	<b>Asbestos Analytical Services</b>	
	Issue No: 010    Issue date: 04 May 2007	
	<b>Unit 5</b> North Harbour Industrial Estate Ayr KA8 8BN	<b>Contact: Mr D Espie</b> Tel: +44 (0)1292 267137 Fax: +44 (0)1292 269025 E-Mail: derekespie@asbestosanalyticalservices.co.uk Website:

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Unit 5 North Harbour Industrial Estate Ayr KA8 8BN	<b>Local contact</b> Mr D Espie	Health and Hygiene  A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



2230  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Asbestos Analytical Services**  
**Issue No: 010 Issue date: 04 May 2007**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	Membrane Filter Method using Phase Contrast Microscopy (PCM) based upon HSG 248	A, B
	Sampling of air for fibre counting	Documented In-House Method No LSD.ASB.02 based upon HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Membrane Filter Method using Phase Contrast Microscopy (PCM) based upon HSG 248	B
	<u>Health and Hygiene</u>		
	Sampling of bulk materials for asbestos identification	Documented In-House Method No LSD.ASB.03	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No LSD.ASB.01 using stereo-microscopy, polarised light microscopy and dispersion staining based upon HSG 248	A
END			