

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Stansted Laboratories Ltd</h3> <p>Issue No: 009 Issue date: 14 June 2011</p>	
	<p>Unit 9 Riverside Industrial Estate 27 Thames Road Barking Essex IG11 0ND</p>	<p>Contact: Mr B Patel Tel: +44 (0)20 8594 5104 Fax: +44 (0)20 8591 8762 E-Mail: sales@stanstedlabs.co.uk Website: www.stanstedlabs.co.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Potable waters and air conditioning waters	<p><u>Microbiological Tests</u></p> <p>Enumeration:</p> <p>Colony count at 22 °C and 37 °C</p>	<p>Documented In-House Method B3.5 based on 'The Microbiology of Drinking Water' (2007) Part 7</p>
Potable waters	<p><i>Pseudomonas aeruginosa</i> <i>Pseudomonas</i> spp</p> <p>Coliforms and <i>Escherichia coli</i></p> <p>Enterococci</p>	<p>Documented In-House Methods B3.9 and B3.10 using CN and CFC agar. B3.10A Species identification using API 20NE</p> <p>Documented In-House Method B3.12 based on 'The Microbiology of Drinking Water' (1009) Part 4, using Colilert 18 Quantitray</p> <p>Documented In-House Method B3.7 based on 'The Microbiology of Drinking Water' (2010) Part 5</p>
Cooling tank waters, tanked waters, showers, air conditioning waters	<p>Detection and enumeration of <i>Legionella</i> spp</p> <p>Identification of: <i>Legionella</i> spp <i>Legionella pneumophila</i> sg 1 <i>Legionella pneumophila</i> sg 2-14</p>	<p>Documented In-House Method B3.8 based on BS 6068 (Parts 4-12), 1998</p>



2266
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Stansted Laboratories Ltd
Issue No: 009 Issue date: 14 June 2011

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Swabs and Contact Plates	<u>Microbiological Tests</u> (cont'd) Total Viable Count at 22 °C Yeasts and Moulds at 25 °C	Documented In-House Method B3.13 1) using TSA/MEA 2) incubation and counting of TSA/MEA plates supplied by Customer
	END	