


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	<h3>AMS Management (GB) LLP</h3> <p>Issue No: 003 Issue date: 28 March 2011</p>	
	<p>Unit 1, 9 Cannon Lane Tonbridge Kent TN9 1PP</p>	<p>Contact: Mr R Steele Tel: +44 (0)1732 368359 Fax: +44 (0)1732 368361 E-Mail: rob.steele@ams-holdings.com Website: www.ams-holdings.com</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Unit 1 9 Cannon Lane Tonbridge Kent TN9 1PP</p> <p>Local contact Gary Wyatt Tel: +44 (0)1732368359 Fax: +44 (0)1732368361 Email: gary.wyatt@ams-holdings.com Website: www.ams-holdings.com</p>	Health and Hygiene	A
<p>Address 16-17 Magnet Point Magnet Road West Thurrock Essex RM20 4DR</p> <p>Local contact Alan Kane Tel: +44 (0)1375 370033 Fax: +44 (0)1375 370025 Email: alan.kane@ams-holdings.com Website: www.ams-holdings.com</p>	Health and Hygiene	B

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	C - Tonbridge



2343

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

AMS Management (GB) LLP
Issue No: 003 Issue date: 28 March 2011

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	A, B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 Documented In-House Procedure AMS LT 001 using stereo-microscopy, polarised light microscopy and dispersion staining	
	Sampling of bulk materials for asbestos identification	Documented In-House Procedure	C
END			