


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS TESTING 2343</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>AMS Management (GB) LLP</h3> <p>Issue No: 002    Issue date: 07 June 2007</p>	
	<p>Unit 1 9 Cannon Lane Tonbridge Kent TN9 1PP</p>	<p>Contact: Mr R Steele Tel: +44 (0)1732 368359 Fax: +44 (0)1732 368361 E-Mail: rob.steele@ams-holdings.com Website: www.ams-holdings.com</p>
<p>Testing performed at the above address only</p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<p><u>Health and Hygiene</u></p> <p>Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite</p>	<p>Health and Safety Executive Asbestos: The Analysts' Guide for Sampling, Analysis and Clearance Procedures (HSG 248)</p> <p>HSG 248:February 2005 by Documented In-House Procedure AMS LT 001 using stereo-microscopy, polarised light microscopy and dispersion staining</p>
	END	