

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Polycom Inc</h3>	
	<p>Issue No: 017 Issue date: 24 November 2011</p>	<p>Singleton Court Business Centre Wonastow Road Industrial Estate (West) Monmouth Monmouthshire NP25 5JA</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>Computers and Peripherals Electronic Products: Digital IT Equipment</p> <p>Office Equipment: Electrical</p>	<p>1 EMC TESTS</p> <p>1.1 Conducted Emissions</p> <p>150 kHz to 30 MHz</p>	<p>CISPR 22:2005 Amd A1 and A2 (Edition 5.2) CISPR 22:2008 (Edition 6.0) EN 55022:2006 including A1:2007 EN 55022:1998 including A1:1995 and A2:1997</p> <p>CFR 47:Part 15B:</p> <p>ANSI C63.4:2003 ANSI C63.4:2009</p> <p>VCCI V-3:2010 VCCI V-3:2011</p> <p>CAN/CSA-CEI/IEC CISPR 22:2002 CAN/CSA-CISPR 22-10</p>
	<p>1.2 Conducted Emissions</p> <p>Signal & Telecom Ports 150 kHz to 30 MHz PSTN Line Shielded Cables ISDN BRI ISDN PRI CAT 5 / 5E 10/100 LAN CAT 5E 1000 LAN</p>	<p>CISPR 22:2005 Amd A1 and A2 (Edition 5.2) CISPR 22:2008 (Edition 6.0) EN 55022:2006 including A1:2007</p> <p>VCCI V-3:2010 VCCI V-3:2011</p>



2389

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Polycom Inc

Issue No: 017 Issue date: 24 November 2011

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
As listed on Page 1	<p>1 EMC TESTS (cont'd)</p> <p>1.3 Radiated Emissions E-Field: 30 MHz to 18 GHz</p>	<p>CISPR 22:2005+A1+A2 (Edition 5.2) CISPR 22:2008 (Edition 6.0) EN 55022:2006 including A1:2007 EN 55022:1998 including A1:1995 and A2:1997</p> <p>CFR 47:Part 15B:</p> <p>ANSI C63.4:2003 ANSI C63.4:2009</p> <p>VCCI V-3:2010 VCCI V-3:2011</p> <p>CAN/CSA-CEI/IEC CISPR 22:2002 CAN/CSA-CISPR 22-10</p>
	<p>1.4 Electrostatic Discharge 0.25 to 16 kV</p>	<p>EN 61000-4-2:1995 EN 61000-4-2:1995 inc Amendment A1:1998 EN 61000-4-2:1995 inc Amendment A1:1998 and A2:2001 EN 61000-4-2 :2009</p>
	<p>1.5 Radiated Susceptibility 80 MHz to 1GHz 6 V/m</p>	<p>EN 61000-4-3:1996 EN 61000-4-3:2006</p>
	<p>1.6 Fast Transient/Burst Immunity: 0.25 kV to 4.0 kV Positive and Negative Polarity 5 ns rise time, 10 ns duration 15 ms burst duration</p>	<p>EN 61000-4-4:1995 EN 61000-4-4:1995 + Amendment A1:2001 EN 61000-4-4:2004</p>



2389

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Polycom Inc

Issue No: 017 Issue date: 24 November 2011

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
As listed on Page 1	<p>1 EMC TESTS (cont'd)</p> <p>1.7 Surge Immunity Waveforms: 0.2 kV to 6 kV 1.2/50 (8/20) μs Common mode Differential mode 10/700 μs 4kV</p>	<p>EN 61000-4-5:1995 including Amendment A1:2001 EN 61000-4-5:2006 ITU K22</p>
	<p>1.8 Conducted RF Susceptibility 150 kHz to 80 MHz 6 V rms</p>	<p>EN 61000-4-6:1996 EN 61000-4-6:2007 EN 61000-4-6 :2009</p>
	<p>1.9 AC Power Ports Voltage Dips, Interruptions and Fluctuations (Immunity): Voltage Dips: up to 100% Voltage Reductions: up to 100% Voltage Fluctuations: up to 100%</p>	<p>EN 61000-4-11:1994 EN 61000-4-11:1994 including Amendment A1:2001 EN 61000-4-11:2004</p>
	<p>1.10 Generic and Product Standards</p> <p>These Generic and Product specific tests are included in this schedule, but limited to those referred basic standards that are explicitly listed in Sections 1.1 to 1.9.</p>	<p>EN 300 386 v.1.3.1 EN 300 386 v.1.3.3 EN 300 386 v.1.4.1 EN 55024:1998 including Amendment A1:2001 & A2:2003 EN 50082-1:1997 EN 61000-6-1:2005 EN 61000-6-1:2007 CISPR 24 1997, Amendment A1:2001 & A2:2003 ICES-003:2004</p>



2389

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Polycom Inc

Issue No: 017 Issue date: 24 November 2011

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
	<p>EMC FACILITIES</p> <p>Screened Anechoic Room Ferrite lined walls and ceiling NSA compliant alternative test site (30MHz to 1GHz) Svswr compliant alternative test site (1GHz to 18GHz) Total Internal Space: 9.5 m x 6 m x 6 m high Variable Single Phase: 50 & 60 Hz, 0 to 230V, 13A supply</p> <p>Screened Anechoic Room Ferrite lined walls and ceiling EN61000-4-6 Uniform Field to 6GHz Total Internal Space: 7.2 m x 3.4 m x 2.7m high Fixed Single Phase: 50 Hz, 230V, 13A Supply</p> <p>Screened Room Total Internal Space: 2.9 m x 2.4 m x 2.7m high Variable Single Phase: 50 Hz, 0 to 230V, 13A supply</p> <p>Ground Plane 3.5m x 2.4m</p>	
	END	