


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Young Calibration Limited	
	Issue No: 004	Issue date: 14 December 2005
	Unit 4, Ham Business Centre Brighton Road Shoreham by Sea West Sussex BN43 6RE	Contact: Mr A Young Tel: +44 (0)1273 455572 Fax: +44 (0)1273 454120 E-Mail: ay@youngcalibration.co.uk Website: www.youngcalibration.co.uk
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ENGINE RADIATORS	<u>Heat rejection tests</u> Flow 10 to 300 L/min Heat rejection 5 to 200 kW	PM53I (In-House) Documented Methods in accordance with customer specifications
END		