


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Asbestos Survey & Management Limited	
	Issue No: 007 Issue date: 07 June 2010	
	Suite 3 New Media Centre 36-42 New Tythe Street Long Eaton Nottingham NG10 2DL	Contact: Mr S Janssen Tel: +44 (0)115 946 2688 Fax: +44 (0)115 946 2689 E-Mail: asm_ltd@btconnect.com Website: www.asmltd.biz

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Suite 3 New Media Centre 36-42 New Tythe Street Long Eaton Nottingham NG10 2DL	Local contact Mr S Janssen	Health and Hygiene Asbestos
		A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Contact: Mr S Janssen	Health and Hygiene Asbestos
		B



2544

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Asbestos Survey & Management Limited

Issue No: 007 Issue date: 07 June 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248: February 2005 by Documented In-House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining	A
ASBESTOS FIBRES IN AIR	Sampling of bulk materials suspected to contain asbestos	Documented In-House Procedure	B, C
	Fibre counting	HSG 248 (November 1995) by Documented In-House Procedure using phase contrast microscopy (PCM)	A, B
	Sampling of air for fibre counting	HSG248: February 2005 (Documented In-House Procedure)	B
	4 Stage Clearance Process	HSG248: February 2005 (Documented In-House Procedure) Membrane Filter Method using Phase Contrast Microscopy (PCM)	B
END			