


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;"><b>2569</b></p> <p style="text-align: center;">Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3 style="margin: 0;">Lucion Environmental Ltd</h3> <p style="margin: 0;"><b>Issue No: 008    Issue date: 24 September 2010</b></p>	
	<p><b>Unit 7</b> Halifax Court Dunston Gateshead Newcastle upon Tyne NE11 9JT</p>	<p><b>Contact: Dr P Morton</b> Tel: +44 (0)191 461 8999 Fax: E-Mail: <a href="mailto:patrickmorton@lucion.co.uk">patrickmorton@lucion.co.uk</a> Website: <a href="http://www.lucion.co.uk">www.lucion.co.uk</a></p>

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> Unit 7 Halifax Court Dunston Gateshead Newcastle upon Tyne NE11 9JT</p> <p style="text-align: right;"><b>Local contact</b> Dr P Morton</p>	<p>Health and Hygiene Asbestos</p> <p><b>Support functions:</b> Quality System Quality Audit Administration</p>	A
<p><b>Address:</b> Suite F24 James House Newport Road Albrighton WV7 3FA</p> <p style="text-align: right;"><b>Local contact</b> Dr P Morton</p>	<p>Health and Hygiene Asbestos</p>	C
<p><b>Address:</b> Unit 2 The Office Building Gatwick Road Manor Royal Crawley RH10 9RZ</p> <p style="text-align: right;"><b>Local contact</b> Dr P Morton</p>	<p>Health and Hygiene Asbestos</p>	D
<p><b>Address:</b> Unit 28 Lisburn Enterprise Park Ballinderry Road Lisburn County Antrim BT28 2BP</p> <p style="text-align: right;"><b>Local contact</b> Dr P Morton</p>	<p>Health and Hygiene Asbestos</p>	E



2569

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Lucion Environmental Ltd**

**Issue No: 008 Issue date: 24 September 2010**

Testing performed by the Organisation at the locations specified

**Site activities performed away from the locations listed above:**

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene Asbestos	B



2569

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Lucion Environmental Ltd**

**Issue No: 008 Issue date: 24 September 2010**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248) and Methods for the Determination of Hazardous Substances in the series MDHS	
	Fibre counting	HSG248: February 2005 by Documented In-House Procedure using phase contrast microscopy (PCM)	A, B, C, D, E
	Sampling of air for fibre counting	HSG248:February 2005 (Documented In-House Procedure)Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B, C, D, E
	4 Stage Clearance Process	HSG248:February 2005 (Documented In-House Procedure)Membrane Filter Method using Phase Contrast Microscopy (PCM)	B, C, D, E
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining	A
ASBESTOS IN SOILS (fibre screening and identification)	Qualitative Asbestos Fibre Screening and Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining based upon HSG 248 (February 2005)	A
END			