

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Enquin Environmental Limited	
	Issue No: 006	Issue date: 01 November 2010
	1 Ringside Business Centre Heol Y Rhosog Wentloog Cardiff CF3 2EW	Contact: Rachel Lewis Tel: +44 (0)29 2053 5090 Fax: +44 (0)29 2077 8633 E-Mail: asbestos@enquin.co.uk Website: www.enquin.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 1 Ringside Business Centre Heol Y Rhosog Wentloog Cardiff CF3 2EW Local contact Rachel Lewis Tel: +44 (0)29 2053 5090 Fax: +44 (0)29 2077 8633 Email: asbestos@enquin.co.uk Website: www.enquin.co.uk	Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



2604

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Enquin Environmental Limited
Issue No: 006 Issue date: 01 November 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure in the Procedures Manual at Section 2)	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure, Quality Control Manual, Section 3, using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for asbestos identification	Documented In-House Method - Quality Control Manual, Section 3, based upon HSG 248 and HSE HSG264	B
END			