





2620

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Stephenson's Analytical Services Limited**

**Issue No: 011 Issue date: 13 May 2011**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	HSG 248:February 2005 Membrane Filter Method - documented in-house methods TP9	A, B, C
	Fibre counting	HSG 248:February 2005 Membrane Filter Method using Phase contrast microscopy (PCM) - documented in-house method TP9	A, B, C
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure) TP9	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure No SAS LSD-001 using stereo-microscopy, polarised light microscopy and dispersion staining	A, B
END			