

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>2650</p> <p>Accredited to ISO/IEC 17025:2005</p>	National Grid Plc	
	Issue No: 008 Issue date: 22 November 2009	
	National Grid House Warwick Technology Park Gallows Hill Warwick CV34 6DA	Contact: Mr H Kinch Tel: +44 (0)1926 656197 Fax: E-Mail: hugh.kinch@uk.ngrid.com Website: www.nationalgrid.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory location:

Location details	Activity	Location code
Address National Grid House Warwick Technology Park Gallows hill Warwick CV34 6DA	Local contact Mr Hugh Kinch Tel: +44 (0)1926 656197 Fax: Email: hugh.kinch@uk.ngrid.com Website: www.nationalgrid.com	Support Functions: Quality System Quality Audit Administration Sampling and Testing: Stack emissions Testing
		A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer sites requiring Stack Emissions Testing	Stack emissions Testing	B



2650

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

National Grid Plc

Issue No: 008 Issue date: 22 November 2009

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere	<u>Sampling and On-Line Analysis</u>	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007	
	Carbon monoxide*	BS EN 15058:2006 (National Grid PPC Emission Monitoring Procedure - NDIR analyser)	B
	Oxygen*	BS EN 14789:2005 (National Grid PPC Emission Monitoring Procedure - Paramagnetic analyser)	B
	Oxides of Nitrogen *	BS EN 14792:2005 (National Grid PPC Emission Monitoring Procedure - Chemiluminescence analyser)	B

END

* - The scale range of the analyser used for this test must be that detailed on its current MCERTS certificate or a range validated by the organisation to meet MCERTS requirements.