


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	PKL Consultants Limited	
	Issue No: 005 Issue date: 24 June 2010	
	Second Floor Offices Victoria House Croft Street Widnes Cheshire WA8 0NQ	Contact: Mr S Longworth Tel: +44 (0)151 422 0330 Fax: +44 (0)151 420 0034 E-Mail: SeanLongworth@pklconsultants.co.uk Website:

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Second Floor Offices Victoria House Croft Street Widnes Cheshire WA8 0NQ Local contact Mr P Longworth Tel: +44 (0)151-422 0330 Fax: +44 (0)151-420 0034 Email: patricklongworth@pklconsultants.co.uk	Asbestos Identification	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Asbestos	B
Mobile Laboratory	Asbestos	C



2658

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

PKL Consultants Limited
Issue No: 005 Issue date: 24 June 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	A
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining	
ASBESTOS FIBRES IN AIR	Sampling of bulk materials suspected to contain asbestos	Documented In-House Procedure	B
	Fibre Counting	HSG 248:February 2005 (Documented In-House Procedure)	B
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	A, B, C
	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	B

END