


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>Accredited to ISO/IEC 17025:2005</b>	<b>Fourstage Environmental Ltd</b>	
	<b>Issue No: 009</b>	<b>Issue date: 11 February 2010</b>
	<b>15 Market Street Castle Donington Derbyshire DE74 2JB</b>	<b>Contact: Mr D Higbed Tel: +44 (0)1332 812549 Fax: +44 (0)1332 814908 E-Mail: fourstage@hotmail.com Website:</b>

**Testing performed by the Organisation at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 15 Market Street Castle Donington Derbyshire DE74 2JB  <b>Local contact</b> Mr D Higbed  Tel: +44 (0)1332 812549 Fax: +44 (0)1332 814908 Email: fourstage@hotmail.com	Health and Hygiene	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



2671

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Fourstage Environmental Ltd**

**Issue No: 009 Issue date: 11 February 2010**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure DOC/002)	A, B, C
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure DOC/002) Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure DOC/002) Membrane Filter Method using Phase Contrast Microscopy (PCM)	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 (Documented In-House Procedure DOC/101) using stereo-microscopy, polarised light microscopy and dispersion staining	A
END			