

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	Micron Analytical Ltd	
	Issue No: 003	Issue date: 17 November 2008
	Unit 6 Exmoor Avenue Skippingdale Industrial Estate Scunthorpe DN15 8NJ	Contact: Miss L Whitehead Tel: +44 (0)1724 859984 Fax: +44 (0)1724 847903 E-Mail: lucy@micronanalytical.co.uk Website: www.micronanalytical.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Unit 6 Exmoor Avenue Skippingdale Industrial Estate Scunthorpe DN15 8NJ	Local contact Miss L Whitehead Tel: +44 (0)1724 859984 Fax: +44 (0)1724 847903 Email: lucy@microanalytical.co.uk	<u>Support Functions</u> Quality System Quality Audit Administration <u>Sampling and Testing</u> Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



4007

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Micron Analytical Ltd

Issue No: 003 Issue date: 17 November 2008

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining	A
ASBESTOS FIBRES IN AIR	Sampling of bulk materials suspected to contain asbestos	Documented In-House Procedure	B
	Fibre counting	HSG248:February 2005 (Documented In-House Procedure)	A, B
	Sampling of air for fibre counting	HSG248:February 2005 (Documented In-House Procedure)	B
	4 Stage Clearance Process	HSG248:February 2005 (Documented In-House Procedure)	B
END			