


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	Westfield Caledonian Ltd	
	Issue No:005 Issue date: 23 August 2010	
	4 Mollins Court Cumbernauld G68 9HP	Contact: Mr J Bowman Tel: +44 (0)1236 786300 Fax: +44 (0)1236 786301 E-Mail: jimbowman@west-cal.co.uk Website: www.west-cal.co.uk
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS – potable and non-potable	Enumeration of: Total Viable Count at 22°C and 37°C	In-House Documented Methods based on the Microbiology of Drinking Water (MDW) 1) LO 001 based on MDW, Part 7, (2007) 2) LO 002 (extended dilution method) based on MDW, Part 7, (2007)
Endoscope Rinse waters	Total Aerobic Colony Count	LO 040, in-house method using membrane filtration onto TSA based on HTM 2030, section 9.223
WATERS – potable and non-potable	Coliform bacteria <i>Escherichia coli</i> <i>Pseudomonas aeruginosa</i> Enterococci Sulphite Reducing Clostridia Legionella pneumophila 1, 2-14 and Legionella species	LO 004 based on MDW, Part 4B (2009) LO 004 based on MDW, Part 4B (2009) LO 005 based on MDW, Part 8, (2010) LO 028 based on MDW, Part 5, (2010) LO 038 based on MDW, Part 6, (2010) LO 009 based on ISO 11731: 1998



4042

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Westfield Caledonian Ltd

Issue No: 005 Issue date: 23 August 2010

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ENVIRONMENTAL AIR AND SURFACE CONTACT PLATES	Aerobic bacteria	In-House Method LO 003
	Moulds (fungi)	In-House Method LO 003
END		