


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>4164</p> <p>Accredited to ISO/IEC 17025:2005</p>	Yara UK Ltd Incorporating Yara Analytical Services and Lancrop Laboratories	
	Issue No: 003 Issue date: 26 October 2011	
	Yara UK Ltd Incorporating Yara Analytical Services and Lancrop Laboratories Manor Place Wellington Road The Industrial Estate Pocklington York YO42 1DN	Contact: Mr A Dawson Tel: +44 (0)1759 302545 Fax: +44 (0)1759 303650 E-Mail: adrian.dawson@yara.com Website: www.yara.com
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
SOILS	<u>Chemical Tests</u>	
	pH	Documented In-house methods - P03.01 (Macauley method using CaCl ₂ as extractant) P18 (Irish Method using Water as extractant) L03 (Buffer Method using SMP (Shoemaker-McLean-Pratt) buffer as extractant) - all using pH meter
	Magnesium Potassium	Documented In-house method M04.01 using a Morgan's extraction and ICP-OES
	Phosphorus	Documented In-house method M04.02 using a Morgan's extraction and discrete colorimetric analysis
BIOTA/VEGETATION/PLANT TISSUE	Organic Matter	Documented In-house method O04 by loss on ignition at 500 °C
	Total N	Documented In-house method (based on Dumas method) N11 using LECO CNS analyser
	END	