


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS TESTING 4224</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p><b>ITS Testing Services (UK) Limited</b> <b>(Ellesmere Port Laboratory)</b></p> <p>Issue No: 004    Issue date: 08 July 2010</p>	
	<p>Rossliffe Road Rossmore Trading Estate Ellesmere Port South Wirral Merseyside CH65 3BS</p>	<p>Contact: Mr Tony Kavanagh Tel: +44 (0)151 355 2005 Fax: +44 (0)151 355 2006 E-Mail: Tony.Kavanagh@Intertek.com Website: www.intertek-cb.com</p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM PRODUCTS, LUBE OILS	<u>Chemical and Physical Tests</u>	Flexible scope enabling new versions of existing accredited standard test methods to be introduced in accordance with documented Intertek corporate procedure COR-PR-14 (Management of Change)
	Elemental Analysis C, H, N	ASTM D5291
	Density	IP 365
	Determination of gross and net heat of combustion	IP 12 ASTM D240
	Viscosity	IP 71 ASTM D445
	Flash Point (COC)	IP 36
	Sulphur Content	IP 336
	RVP (minivap)	ASTM D5191
	Colour	ASTM D1500
	Pour Point	IP 15 ASTM D97
Water Content	IP 438	



4224

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**ITS Testing Services (UK) Limited**  
**(Ellesmere Port Laboratory)**

**Issue No: 004 Issue date: 08 July 2010**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PETROLEUM PRODUCTS, LUBE OILS (cont'd)	<u>Chemical and Physical Tests</u> (cont'd)	Flexible scope enabling new versions of existing accredited standard test methods to be introduced in accordance with documented Intertek corporate procedure COR-PR-14 (Management of Change)
	Aniline Point	IP 2 ASTM D611
	Refractive Index	ASTM D1218
	Distillation (automatic only)	ASTM D86
LUBE OILS	Apparent viscosity (Cold Cranking Simulator)	ASTM D5293
	Evaporative Loss (Noack)	ASTM D5800
SOLVENTS	Distillation (automatic only)	ASTM D1078
	END	