


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS TESTING 4346</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>RWE npower plc</h3> <p>Issue No: 003 Issue date: 22 November 2010</p>	
	<p>Chemistry & Environment Department Didcot A Power Station Didcot Oxfordshire OX11 7HA</p>	<p>Contact: This laboratory does not accept commercial or external enquiries</p> <p>Tel: Fax: E-Mail: Website: www.rwenpower.com</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Chemistry and Environment Division Didcot A Power Station Didcot Oxfordshire OX11 7HA</p> <p>Local contact This laboratory does not accept commercial or external enquiries</p> <p>Tel: Fax: Email: Website: www.rwenpower.com</p>	<p>Support Functions: Quality System Quality Audit Administration</p> <p>Sampling and Testing: Chemical Tests</p>	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Consented water discharges for RWE NPower PLC, Didcot	Environmental Sampling	B



4346

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

RWE npower plc

Issue No: 003 **Issue date:** 22 November 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS - River water and site industrial effluent	<u>Sampling for chemical analysis conducted in-house</u>	Documented In-House Methods	B
	<u>Chemical Analysis</u>		
	Total Chlorine (0 - 5 mg/l)	Documented In-House Method to BS EN ISO 7393-2:2000 (DID-LAB-008)	A
	Oil in Water (0 - 200 mg/l)	Documented In-House Method using Partition Infrared based on standard method 5520C for examination of water and waste water (DID-LAB-009)	A
	pH (0 - 14)	Documented In-House Method to BS 6068-2.5:1995 (DID-LAB-006)	A
	Suspended solids (≥ 2 mg/l)	Documented In-House Method to BS EN 872:2005 (DID-LAB-010)	A
	Chloride Sulphate	Documented In-House Method to BS EN ISO 10304-1:1995 (DID-LAB-007)	A
END			