


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

| | | |
|---|---|--|
|  <p>4365</p> <p>Accredited to ISO/IEC 17025:2005</p> | <h3>Smith & Nephew Orthopaedics</h3> <p>Issue No: 001 Issue date: 20 November 2009</p> | |
| | <p>Aurora House Harrison Way Leamington Spa Warwickshire CV31 3HL</p> | <p>Contact: Mr Azad Hussain Tel: +44 (0)1926 482376 Fax: +44 (0)1926 482499 E-Mail: Azad.Hussain@smith-nephew.com Website:</p> |
| <p>Testing performed at the above address only</p> | | |

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|---|--|---|
| <p>IMPLANTS FOR SURGERY -TOTAL HIP-JOINT PROSTHESES</p> | <p><u>Physical testing</u></p> <p><u>Wear of total hip-joint prostheses:</u></p> <ul style="list-style-type: none"> - Wear loss following hip-joint simulator tests - Friction testing | <p>ISO 14242-1: 2002 ISO 14242-2 :2000 ISO 14242-3 :2009</p> <p><u>In-House Methods</u> (based on peer-reviewed methods from journals) WI 120, Rev 1, 30 Sep 2009</p> |
| | <p>END</p> | |