


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS TESTING</b> 4522</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p><b>Theatre &amp; Endoscopy Service Technicians Ltd (T.E.S.T. Ltd)</b></p> <p>Issue No: 001    Issue date: 27 May 2011</p>	
	<p><b>Test House</b> Unit 8 Whieldon Road Industrial Estate Stoke-on-Trent Staffordshire ST4 4JP</p>	<p><b>Contact: Mr T Rowley</b> Tel: +44 (0)1782 572400 Fax: +44 (0)1782 414839 E-Mail: timrowley@testmedical.co.uk Website: www.testmedical.co.uk</p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p><b>ENVIRONMENTAL SAMPLES</b> Tubing samples related to Washer Disinfector Units used in Healthcare Institutions</p>	<p><u>Microbiological Tests</u> Evaluation of Biofilm levels as a measure of Disinfection Efficacy</p> <p>Evaluation of Bacterial Contamination levels as a measure of Disinfection Efficacy</p>	<p>Documented In-House methods:</p> <p>SOP 106 assessment of <i>Pseudomonas aeruginosa</i> biofilm loading on PTFE tubing before and after a Washer Disinfector Cycle at customer premises, based on HTM 2030 1997, Section 17</p> <p>SOP 107 assessment of contamination levels of <i>Pseudomonas aeruginosa</i>, <i>Escherichia coli</i> and <i>Staphylococcus aureus</i> levels on PTFE tubing or stainless steel discs (surrogate devices) before and after a Washer Disinfector Cycle at customer premises, based on HTM 2030 1997, Section 11</p>
<p><b>WATERS</b> RO Water, Endoscopy rinse waters and feed/source water used with Washer Disinfector Units in Healthcare Institutions</p>	<p><u>Microbiological Tests</u> Bacterial Endotoxins (LAL Test)</p>	<p>Documented In-House methods:</p> <p>SOP 108 chromogenic assay using Charles River portable Test system (PTS) Kinetic Reader and Endosafe cartridges, based on HTM 2030 1997, Section 9.201 to 9.212</p>



4522

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Theatre & Endoscopy Service Technicians Ltd**  
**(T.E.S.T. Ltd)**

**Issue No: 001 Issue date: 27 May 2011**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS RO Water, Endoscopy rinse waters and feed/source water used with Washer Disinfectant Units in Healthcare Institutions (cont'd)	<u>Microbiological Tests</u> (cont'd)  Total Viable Count	Documented In-House methods:  SOP 100 using membrane filtration on Tryptone Soya Agar incubated at $35 \pm 2$ °C for 2 days, based on HTM 2030 1997, Section 9.213 to 9.226
	END	