


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS TESTING 4619</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Covance Laboratories Limited</h3> <p>Issue No: 001 Issue date: 18 May 2011</p>	
	<p>Otley Road Harrogate HG3 1PY</p>	<p>Contact: Ms L Gillbanks Tel: +44 (0)1423 500011 Fax: +44 (0)1423 569595 E-Mail: linzi.gillbanks@covance.com Website: www.covance.com</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
CIGARETTES Cigarette smoke	<p><u>Chemical Tests</u></p> <p>Determination of total particulate matter and puff number</p> <p>Determination of nicotine free dry particulate matter</p> <p>Determination of water in smoke condensates</p> <p>Determination of nicotine in smoke condensates</p>	<p>In-house method GMT 05-070 based on ISO 4387:2000 using Borgwaldt RM 200 rotary smoking machine</p> <p>In-house method GMT 18-005 by calculation</p> <p>In-house method GMT CLE 8231-112-A based on ISO 10362-1:1999 by GC</p> <p>In-house method GMT CLE 8231-112-A based on ISO 10315:2000 by GC</p>
	END	