


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Littlebrook Calibration Services Ltd	
	Issue No: 003 Issue date: 20 June 2011	
	5 Optima Park Thomas Road (off Thames Road) Crayford Kent DA1 4QX	Contact: Mr E A Thompson Tel: +44 (0)1322-556111 Fax: +44 (0)1322-520400 E-Mail: lcmdirector@17025calibrations.com Website: www.17025calibrations.com
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HOSES, PIPES and TUBES PRESSURE VESSELS PRESSURE RELIEF VALVES	Pneumatic pressure testing Up to 70 bar Hydraulic pressure testing Up to 600 bar	Documented In-house Method PROC 31 Issue 1 Version 3 dated 08/11/2010
END		